Mnemonic's

AC0VXG	mains supply 0V X-ray generator
AC230VL1	mains supply 230V AC phase 1
AC230VL2	mains supply 230V AC phase 2
GND12V	ground +12V supply
MNEMOF	mains power emergency off
MNON/	mains on
PO12V	+12V supply
PWONNG	DC supply relay power on -
PWONPO	DC supply relay power on +
RDMNON/	ready mains power on
RESETSW/	system reset
SIPHIO	single phase intijicotion

Wall junction box
MEX
Mains supply
3~400V
50/60Hz
F1 25A
K1
K2
R1 47R
R2 47R
R3 47R
EQX1101 (E2QX1101)
EQX1102 (E2QX1102)
EQX1103 (E2QX1103)
EQ200 X12 (E2Q200 X12)
EQ200 X11 (E2Q200 X11)
Discharging circuit
X3201
F2 10A
F12 10A
X23
X22
NX21
K11
Output 3
Imaging
Output 2
II/TV
GECO
SYSCO
Output 1
Geometry
F11 5A
F3 5A
K11
EM off
EN
X3:1
X3:3
T1
230V~
X3:2
H1 gr
low voltage control
100
Power on circuit
S1
V1
X1:6
X1:7
X1:2
X1:1
X2:1
X2:2
X1:9
X1:5
X1:4
T1 X2402
T2 X2403
RC1
X1102
WG 51
WG 57
EZ X13:3
EZ X13:2
EZ X13:1
EZ X14:1
EZ X14:2
EZ X15:1
EZ X15:2
EZ X47:6
EZ X47:7
EZ X47:2
EZ X47:1
EZ X47:9
EZ X47:5
EZ X47:4
EZ X4:1
EZ X4:2
GND 12V
RDMNON/
MNON/
PO26V
GND26V
PWONNG
SIPHIO
PWONPO
EZ X46:15
EZ X46:14
EZ X46:6
EZ X20:1
EZ X20:2
X10:11
X10:4
X10:3
X10:1
X10:15
Logic
off
on
CB100 Cockpit
YM
M1 RoCo
GS1:1
M cooler
H.V. generator
G200 X2
Oil circulation pump
H.V. generator
RoCo LS
RoCo HS
WG
Tube extension (option)
EZ119 mA control
EZ102 low voltage supply
EWN second low voltage supply (option)
Adaptation 24V-/230V~
EZ130 kV control
see Z1 :H.V. generator R/F
Power supply

A2/A3 02-04-30 Schr./Kö

© Philips Medical Systems